2 - Specifications

VOLTAGE

Five DC and AC Ranges:

15, 30, 150, 300 & 600 Volts DC & Volts rms.

Remote Voltage Sensing:

- Differential input 110 dB CMRR.
- Maximum of 30 Volts peak, volts low terminal to amps output terminal.

CURRENT

Five DC and AC Ranges:

• 2, 5, 10, 20 and 50 Amps DC and Amps rms.

METER IMPEDANCE

Voltage, DC or AC:

3 Megohm load on each voltage range.

Current, DC or AC Ranges:

- 0.011 Ohm shunt resistance in the 2, 5, 10 and 20 Amp ranges.
- 0.001 Ohm shunt resistance in the 50 Amp range.

RESOLUTION

Processing resolution is 16 binary bits.

Voltage Display:

- All ranges $\leq 9.999 \text{ V}$ is $\pm 0.001 \text{ Volt}$.
- Ranges $\ge 10 \text{ V}$ and $\le 99.9 \text{ V}$ is $\pm 0.01 \text{ Volt}$.
- Ranges $\geq 100 \text{ V}$ is $\pm 0.1 \text{ Volt}$.

Current Display:

- All ranges ≤ 9.999 A is ± 0.001 Amp.
- Ranges $\geq 10 \text{ A}$ and $\leq 50 \text{ A}$ is $\pm 0.01 \text{ Amp}$.

Power Display:

- ± 0.01% of active Volts and Amps range
- GPIB Amps, Volts and Watts Same as display resolution.

ISOLATION

1500 Vrms break down from input circuit to chassis (ground).

DISPLAY AUTO ZERO

When the A and V, display indication is less than 0.5 % of range, the displayed value is set to ZERO. Refer to SECTION 5 - SPECIAL FUNCTIONS to disable the AUTO ZERO function.

ANALOG OUTPUTS

This is an optional feature.

Analog AMPS, VOLTS and WATTS output signals are a relative value of approximately 5 Volts DC for full scale on each range selected. These outputs are calibrated to 5.00 Volts DC \pm 0.2% for full scale input on the 10 Amp current range and 150 Volt voltage range (1500 Watt range). Ripple is less than 5 mV.

The outputs are operational amplifiers with an impedance of < 1 Ohm and current capacity of < 4 mA. Each output signal and common is electrically isolated from the measured circuit by 750 Volts continuous and 2500 Volts test breakdown. Leakage current is less than 0.3 μ A at 240 Vrms, 60 Hz.

DATA ACQUISITION

ANALOG

The conversion of true rms to DC is expressed as:

$$V_{\rm rms} \equiv \sqrt{\frac{1}{T}} \int_{0}^{T} v^{2}(t) dt$$

Rms to DC conversion averaging time constant:
 Volts and Amps time constant TC ≈ 60 msec.
 Watts time constant TC ≈ 120 msec.
 Watts = V × I × cos θ (instantaneous)

DIGITAL

- Processing:
 - Integration period = 0.10 second.

 Display update time = 2 readings per second.
- IEEE-488 (GPIB):
 Synchronized = 0.1 sec. per reading
 Non-synchronized is 0.04 to 0.07 sec per reading.

MEASUREMENT ACCURACY

Specified test conditions: Ambient temperature of 72° ± 10°F and power factor of 0.1 to 1.0, lead or lag.

VOLTAGE - DC	\pm (0.1% of reading + 0.2% of range)
VOLTAGE - AC:	1,15
10 Hz to < 20 Hz	± 1.0% of range
20 Hz to < 45 Hz	± (0.2% of reading + 0.3% of range)
45 Hz to < 10 kHz	± (0.1% of reading + 0.2% of range)
10 kHz to 20 kHz	± (0.2% of reading + 0.3% of range)
CURRENT - DC	2, 5, 10, 20 and 50 Amp Ranges
	± (0.1% of reading + 0.2% of range)
CURRENT - AC:	2, 5, 10 and 20 Amp Ranges
10 Hz to < 20 Hz	± 1.0% of range
20 Hz to < 45 Hz	± (0.2% of reading + 0.3% of range)
45 Hz to < 1 kHz	± (0.1% of reading + 0.2% of range)
1 kHz to < 5 kHz	± (0.2% of reading + 0.3% of range)
5 kHz to < 10 kHz	± 1.0% of reading
10 kHz to 20 kHz	± 2.0% of reading
	50 Amp Range
10 Hz to < 1 kHz	same as 20 Amp range
1 kHz to 2 kHz	± 2.0% of range
POWER - DC	\pm (0.1% of reading + 0.3% of VA range)
POWER - AC	± (0.1% of reading + 0.2% of VA range)
POWER FACTOR	± (VA error ± W error)
DISPLAY	Digital display error ± 1 LSB
CREST FACTOR	Exceeds 3:1 (at 50% of range full scale)
TEMPERATURE COEFFICIENT	± 0.01% of range per deg. C. max

ACCURACY CERTIFICATION

All instruments are shipped with a CERTIFICATE of CALIBRATION from MAGTROL Inc. Magtrol policies and procedures comply with MIL-STD-45662A. Measurement standards are traceable to the National Institute of Standards and Technology (NIST).

Instrument calibration every six calendar months is necessary to maintain full compliance with all specifications.

If a one year calibration cycle is used, all accuracy specifications are reduced by 0.1%. After one calendar year, the instrument is considered to be out of calibration.