## SPECIFICATIONS

All specifications are tested in accordance with standard California Instruments test procedures. All specifications apply over the frequency range from 45Hz to 10KHz. Consult the manual addendum for the actual oscillator output frequency.

MODEL

NUMBER:

815T-Freq-. 1-1

815T-Freq-.1-3\* 815T-Freq-.1-2

No. of Phases:

single phase

three phase

two phase

Nominal

Phase Shigt:

Not applicable

± 120 degrees

+ 90 degrees

Phase

Sequence:

Not applicable

ABC sequence

CA sequence

Amplitude

Tracking:

Not applicable

1.0 per cent of

1.0 per cent of

full scale

full scale

Phase Accuracy

(with amplitude potentiometer adjusted from 75% of full output to full output)\*\*

at 25°C:

(with amplitude potentiometer adjusted from 75% of full output to full output) ##

Not applicable

±1.0 degree maximum.

±1.0 degree maximum.

## SPECIFICATIONS COMMON TO ALL MODELS

Frequency

Single Fixed Frequency, 45Hz to 10KHz.

\_ Frequency Accuracy at 25°C:

±0.10 percent.

Frequency Temperature Coefficient:

±.02 percent/°C maximum

Output Amplitude:

5.0 volts rms minimum for single phase versions; U to greater than 5 volts rms for multiphase versions.

Amplitude Stability:

(after one hour warm-up)

±0.25 percent for 24 hours at 25°C

±0.02 percent/°C maximum

Total Harmonic Distortion:

0.2 percent maximum

Input Power:

±25 volts DC at approximately 18 milliamperes available from power source for single phase versions:

±25 volts DC at approximately 55 milliamperes available from power source for three phase versions.

Operating Temperature Range:

0 to 55°C for all versions.

Dimensions:

3-1/2" high x 8" wide x 7" deep for mounting into a California Instruments Solid State Invertron.

Front Panel Finish:

Grey 26440. Federal Standard 595.

\* Order Model 815T-Freq..1-3D for use in the three phase open delta configuration.

\*\* This is a new improved specification. Oscillators manufactured prior to October 1974 were specified at 2% amplitude tracking with the AMPLITUDE control fully clockwise and 2° phase error with the AMPLITUDE control fully clockwise. These early oscillators were not specified as to additional phase and amplitude tracking errors when the AMPLITUDE control was turned counter clockwise from the fully clockwise position.

Step 5.6.4 was added to this manual at the "D" Revision to test for these improved characteristics in oscillators manufactured after October, 1974.