

Fluke 70 Series III Digital Multimeter

Technical Specifications	
DCV	Maximum Voltage: Fluke 70, 73, 21, 75: 600V Fluke 23, 77, 26, 79: 1000V Basic Accuracy: 0.3% Maximum Resolution: 0.1 mV
ACV	Maximum Voltage: Fluke 70, 73, 21, 75: 600V Fluke 23, 77, 26, 79: 1000V Basic Accuracy: Fluke 70, 73, 21, 75, 23, 77: 2% Fluke 26, 79: 1.5% Maximum Resolution: Fluke 70, 73, 21, 75, 23, 77: 1 mV Fluke 26, 79: 0.1mV
DC & AC current	Maximum Amps: 10A continuous (except Fluke 70) Maximum Resolution: Fluke 73, 21, 75, 23, 77: 0.01 mA Fluke 26, 79: 0.001 mA
Resistance	Max Resistance: Fluke 70, 73, 21, 75, 23, 77: 32 M Ω Fluke 26, 79: 40 M Ω Basic Accuracy: Fluke 70, 73, 21, 75, 23, 77: 0.5% Fluke 26, 79: 0.4% Maximum Resolution: Fluke 70, 73, 21, 75, 23, 77: 0.1 Ω Fluke 26, 79: 0.01 Ω
Capacitance	0.01 nF -9,999 μ F (26, 79 only)
Frequency	1.0 Hz to 20 kHz (26, 79 only)