# **Model SR1**

#### RESISTANCE STANDARDS & INSTRUMENTS

- Versatile working standard
- 50 ppm long-term accuracy for most values
- Select from a wide range of values from 0.01  $\Omega$  to 10  $M\Omega$
- Accurate, stable, low cost

# **Standard Resistor**

The Model SR1 Standard Resistor is a laboratory standard of high accuracy and stability. The resistance of most values is initially adjusted to an accuracy of 20 ppm of nominal, with long-term accuracy guaranteed to better than 50 ppm. Other values have initial accuracy from 50 ppm to 200 ppm.

Model SR1 has been constructed to meet today's high standards of performance. The resistance wire used is a modern alloy with excellent stability, an extremely low temperature coefficient over a wide range of temperatures, and very low thermal EMF to copper. A unifilar winding on a flat mica card is used to minimize both series inductance and shunt capacitance. The durable aluminum case provides electrostatic shielding. Gold plated terminals reduce connection errors.

Terminals appear on the top of each unit. They are also brought out the bottom of the case on removable banana plugs. This allows plugging two or more units together in either series or parallel for a wide variety of resistance values. Four-terminal resistance measurements are easily made by using the banana plugs on the bottom of the case as two of the terminals. 0.01 and  $0.1\Omega$  values have four terminals on the top and should only be used in four-wire configurations.

For three-terminal or five-terminal guarded measurements, a binding post on top and a banana plug on the bottom provide connections to the case.





#### STANDARD RESISTOR

# **Specifications**

Standard Values   0.01Ω, 0.1Ω, 1 Ω, 10 Ω, 100 Ω, 1 kΩ, 10 kΩ,   100 kΩ, 1 MΩ, 10 MΩ   Accuracy   See table to the right	Accuracy				Coefficients		Maximum Ratings		
		Initial	Long Term	Cali- bration	Temp.	Power	Power	Current	
	(Ω)	(ppm)	(ppm)	(ppm)	(ppm/°C) (ppm/mW)		(mW)	(mA)	(peak V)
	0.01	200	500	50	60	20	60	2500	
Calibration Conditions 23°C, low power, four-terminal measurement	0.1	100	300	30	40	2.5	250	1600	
	1	20	50	10	15	0.3	1000	1000	
Temperature Coefficient See Table	10	20	50	10	15	0.3	1000	320	
	100	20	50	10	5	0.1	1000	100	
	1 k	20	50	10	5	0.1	1000	32	
Power Coefficient	10 k	20	50	10	5	0.1	1000	10	
See Table	100 k	20	50	10	5	0.1	1000	3.2	
Terminal Resistance	1 M	50	100	20	5	0.1	100	0.3	300
Units with $1\Omega$ or higher resistance: Binding	10 M	50	100	20	5	0.1	10	0.03	300

#### **Initial Accuracy:**

The specifications stated in the TEGAM instrument catalogs and data sheets are intended as acceptance specifications and are guaranteed for 60 days from the date of shipment. They are typically maintained for a much longer period of time.

#### Long-Term Accuracy:

These specifications are guaranteed for the standard warranty period, and are typically maintained for the life of the instrument. Long-term accuracy is implied when not otherwise stated.

#### **Calibration Accuracy:**

Calibration accuracy is the accuracy of TEGAM calibration data relative to the legal units maintained by the U.S. National Institute of Standards and Technology.

Units w posts ad tance value; banana plugs add 2 to 3 m $\!\Omega$  additional resistance

#### **Maximum Ratings** See Table

Breakdown Voltage 1500 V peak to case

## **Calibration Data**


Initial calibration readings are affixed to instrument

#### Dimensions

Height: 2.1" (5.3 cm) Width: 3.8" (9.65 cm) Depth: 2.4" (6.1 cm) Weight: 8 oz. (227 gm) net

### **Standard Equipment**

8234 Instruction Sheet



This data sheet was current when it was produced. However, products are constantly being updated and improved. Because of this some differences may occur between the descriptions herein and the current product. Prices and specifications may be changed without notice.

